



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores, Santo Domingo 10137, República Dominicana

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Chemical, Dimensional, Electrical, Mass, Force, and Weighing Device, Mechanical, Optical, Thermodynamic, Time and Frequency Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

February 11, 2023

Issue Date:

February 11, 2023

Expiration Date:

June 30, 2025

Revision Date:

March 07, 2024

Accreditation No.:

108843

Certificate No.:

L23-110-R1

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
pH Meter/Probe ^{FO}	4 pH to 10 pH	0.009 pH	pH Standard Solutions	PR-CAL-021
Conductivity Meter/Probe ^{FO}	5 μ S/cm	0.62 μ S/cm	Conductivity Standard Solutions	PR-CAL-022
	10 μ S/cm	0.62 μ S/cm		
	25 μ S/cm	0.62 μ S/cm		
	111.3 μ S/cm	0.97 μ S/cm		
	1015 μ S/cm	5.4 μ S/cm		
	1408 μ S/cm	6.9 μ S/cm		
	1413 μ S/cm	6.2 μ S/cm		
	12.85 mS/cm	0.36 mS/cm		
Refractometers ^{FO}	1.355 n	0.000 29 n	Refraction Standard Liquids	PR-CAL-024
	1.420 n	0.000 26 n		
	1.430 n	0.000 26 n		
	1.480 n	0.000 31 n		
	14.94 °Brix	0.15 °Brix		
	55.03 °Brix	0.11 °Brix		
	76.23 °Brix	0.092 °Brix		
Turbidity Meter/Probe ^{FO}	0.04 NTU	0.058 NTU	Turbidity Standard Solutions	PR-CAL-040
	20 NTU	0.63 NTU		
	100 NTU	5.9 NTU		
	200 NTU	6.3 NTU		
	800 NTU	11 NTU		
	1 000 NTU	13 NTU		
	4 000 NTU	47 NTU		



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	
Calipers ^{FO}	0.05 in to 8 in	$(289 + 9.73 \times 10^{-2}L) \mu\text{in}$	Gage Blocks	PR-CAL-032	
	8 in to 12 in	$(287 + 0.35L) \mu\text{in}$			
	12 in to 24 in	$(288 + 0.25L) \mu\text{in}$			
Micrometers ^{FO}	0.05 in to 1 in	$(3.92 + 2.4L) \mu\text{in}$		Master blocks	PR-CAL-033
	1 in to 8 in	$(4.15 + 2.2L) \mu\text{in}$			
Indicators, dial, digital ^{FO}	0.05 in to 6 in	$(119 + 1.9L) \mu\text{in}$			
Rules ^{FO}	0.05 in to 24 in	0.009 in	Master blocks	PR-CAL-035	
Tapes ^{FO}	0.05 in to 300 in	$(0.02 + 5.4 \times 10^{-4}L) \text{in}$			
Pin gages ^{FO}	0.011 in to 1 in	105 μin	Micrometer	PR-CAL-065	

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	
Equipment to measure DC voltage ^{FO}	1 mV to 75 mV	0.025 % of reading + 12 μV	Fluke 724	PR-CAL-026	
	75 mV to 100 mV	0.022 % of reading + 20 μV			
	0.1 V to 10 V	0.020 % of reading + 6.2 mV			
Equipment to output DC voltage ^{FO}	1 mV to 90 mV	0.021 % of reading + 20 μV			
	0.09 V to 30 V	0.021 % of reading + 2 mV			
Equipment to output DC current ^{FO}	1 mA to 24 mA	0.21 % of reading + 2 μA			
Equipment to Measure Resistance ^{FO}	15 Ω to 400 Ω	101 m Ω			
	400 Ω to 1 500 Ω	504 m Ω			
	1 500 Ω to 3 200 Ω	1.0 Ω			
Equipment to output Resistance ^{FO}	0.2 Ω to 400 Ω	101 m Ω			
	400 Ω to 1 500 Ω	504 m Ω			
	1 500 Ω to 3 200 Ω	1.0 Ω			



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type B ^{FO}	600 °C to 800 °C	2.2 °C	Fluke 724 Electrical Simulation of Thermocouple Output	PR-CAL-026
	800 °C to 1 000 °C	1.8 °C		
	1 000 °C to 1 800 °C	1.4 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type E ^{FO}	-200 °C to 0 °C	0.9 °C		
	0 °C to 950 °C	0.7 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J ^{FO}	-200 °C to 0 °C	1.0 °C		
	0 °C to 1 200 °C	0.7 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type K ^{FO}	-200 °C to 0 °C	1.2 °C		
	0 °C to 1 370 °C	0.8 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type L ^{FO}	-200 °C to 0 °C	0.85 °C		
	0 °C to 900 °C	0.7 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type N ^{FO}	-200 °C to 0 °C	1.5 °C		
	0 °C to 1 300 °C	0.9 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type R ^{FO}	-20 °C to 0 °C	2.5 °C		
	0 °C to 500 °C	1.8 °C		
	500 °C to 1 750 °C	1.4 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type S ^{FO}	-20 °C to 0 °C	2.5 °C		
	0 °C to 500 °C	1.8 °C		
	500 °C to 1 750 °C	1.5 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type T ^{FO}	-200 °C to 0 °C	1.2 °C		
	0 °C to 400 °C	0.8 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type U ^{FO}	-200 °C to 0 °C	1.1 °C		
	0 °C to 400 °C	0.75 °C		



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Ni120, 120 Ω ^{FO}	-80 °C to 260 °C	0.20 °C	Fluke 724 Electrical Simulation of RTD Output	PR-CAL-026
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Pt 385, 100 Ω ^{FO}	-200 °C to 800 °C	0.33 °C		
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Pt 392, 100 Ω ^{FO}	-200 °C to 630 °C	0.30 °C		
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Pt 385, 200 Ω ^{FO}	-200 °C to 250 °C	0.20 °C		
	250 °C to 630 °C	0.80 °C		
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Pt 385, 500 Ω ^{FO}	-200 °C to 500 °C	0.30 °C		
	500 °C to 630 °C	0.40 °C		
Temperature Calibration, Indication and Control Equipment used with RTD Indicators / Detectors Type Pt 385, 1 000 Ω ^{FO}	-200 °C to 630 °C	0.20 °C		



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Mass, Force, Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Scales and balances ^{FO}	1 mg to 500 mg	$(7.0 \times 10^{-3} + 4.4 \times 10^{-5}Wt)$ mg	OIML E2 weights	PR-CAL-020
	1 g to 100 g	$(0.1 + 1.5 \times 10^{-3}Wt)$ mg		
	100 g to 200 g	$(0.1 + 1.6 \times 10^{-3}Wt)$ mg		
	200 g to 1 000 g	$(-0.1 + 1.9 \times 10^{-3}Wt)$ mg		
	1 000 g to 5 000 g	$(0.1 + 1.9 \times 10^{-3}Wt)$ mg		
	5 000 g to 10 000 g	$(0.2 + 1.9 \times 10^{-3}Wt)$ mg		
	10 kg to 20 kg	$(0.6 + 1.8 \times 10^{-3}Wt)$ mg		
	20 kg to 40 kg	$(-206 + 1.2 \times 10^{-2}Wt)$ mg	OIML E2, F1, F2 weights	
Scales and weighing devices ^{FO}	40 kg to 300 kg	$(8.8 + 9.3 \times 10^{-2}Wt)$ g	ASTM 6 weights	
	300 kg to 1 600 kg	$(-46.5 + 0.26Wt)$ g		
Mass Weights ASTM Class 2, 3, 4, 5, 6 & 7 OIML Class F1, F2, M1, M2 & M3 NIST Class F ^{FO}	1 mg	8.7 μ g	OIML E2 weight set Mass Comparator	PR-CAL-051
	2 mg	8.7 μ g		
	5 mg	8.7 μ g		
	10 mg	8.7 μ g		
	20 mg	9.0 μ g		
	50 mg	9.4 μ g		
	100 mg	10 μ g		
	200 mg	11 μ g		
	500 mg	12 μ g		
	1 g	0.022 mg		
	2 g	0.033 mg		
	5 g	0.038 mg		
	10 g	0.055 mg		
	20 g	0.080 mg		
	50 g	0.11 mg		
	100 g	0.13 mg		
	200 g	0.22 mg		
	500 g	1 mg		
	1 kg	1.2 mg		
	2 kg	1.6 mg		
	3 kg	1.8 mg		
	5 kg	3.8 mg		
	10 kg	7 mg		



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Mass, Force, Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Mass Weights NIST Class F, ASTM 5, 6 & 7 ^{FO}	20 kg	26 mg	OIML E2 and F1 weights Mass comparator	PR-CAL-051
	25 kg	31 mg		
Force Gauges (Tension and Compression) ^{FO}	10 lbf to 100 lbf	0.079 lbf	Reference Gauge MR04-100	PR-CAL-069
	100 lbf to 500 lbf	0.42 lbf	Reference Gauge MR01-500	

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pressure Gauges, Vacuum Gauges ^{FO}	-14 psig to 0.2 psig	0.023 psig	ADT681 and pneumatic pump	PR-CAL-037
	0.2 psig to 60 psig	0.023 psig		
	60 psig to 180 psig	0.039 psig		
	180 psig to 240 psig	0.054 psig		
	240 psig to 300 psig	0.064 psig		
Differential Pressure Gauges, Pressure Gauges, ^{FO}	0.3 hPa to 400 hPa	0.1 hPa	Testo 526-2 and pneumatic pump	PR-CAL-037
	400 hPa to 800 hPa	0.13 hPa		
	800 hPa to 1 200 hPa	0.18 hPa		
	1 200 hPa to 1 600 hPa	0.25 hPa		
	1 600 hPa to 2 000 hPa	0.39 hPa		
Differential Pressure Gauges Pressure Gauges ^{FO}	-10 in H ₂ O to + 10 in H ₂ O	0.005 6 in H ₂ O	ADT681 and pneumatic pump	PR-CAL-037
Anemometers, Air Velocity Meters ^{FO}	0.4 m/s to 30 m/s	2.1 % of reading + 0.05 m/s	Reference Air Velocity Meter Traceable 4091	PR-CAL-053
Fume Hoods, Laminar Flow Hoods, Biosafety Cabinets, Air velocity only ^{FO}	0.4 m/s to 30 m/s	2.1 % of reading + 0.05 m/s	Reference Air Velocity Meter Traceable 4091	PR-CAL-056
Pipettes, Burettes, Dispensers ^{FO}	0.25 μ L to 20 μ L	0.08 μ L	Gravimetric method reference to mass balances and OIML Class E2 mass standards, Analytical Balance.	PR-CAL-049
	20 μ L to 100 μ L	0.094 μ L		
	100 μ L to 200 μ L	0.12 μ L		
	200 μ L to 500 μ L	0.21 μ L		
	500 μ L to 1 000 μ L	0.35 μ L		
	1 000 μ L to 2 500 μ L	0.8 μ L		



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pipettes, Burettes, Dispensers ^{FO}	2 500 μ L to 5 000 μ L	1.6 μ L	Gravimetric method reference to mass balances and OIML Class E2 mass standards, Analytical Balance.	PR-CAL-049
	5 000 μ L to 10 000 μ L	2.5 μ L		
	10 000 μ L to 20 000 μ L	5.1 μ L		
	20 000 μ L to 50 000 μ L	16 μ L		
	50 000 μ L to 100 000 μ L	28 μ L		
Tachometer – Optical and Mechanical Rotational measurement Centrifuges ^{FO}	5 rpm to 99.999 rpm	$(6.5 \times 10^{-4} + 9.4 \times 10^{-5}R)$ rpm	Comparison to Standard Tachometer PLT200	PR-CAL-046
	100 rpm to 999.99 rpm	$(4.5 \times 10^{-3} + 9.6 \times 10^{-5}R)$ rpm		
	1 000 rpm to 9 999.9	$(3.2 \times 10^{-2} + 9.7 \times 10^{-5}R)$ rpm		
	10 000 rpm to 99 999	$(1.3 + 2.3 \times 10^{-6}R)$ rpm		
	100 000 rpm to 200 000 rpm	$(11 + 4.0 \times 10^{-6}R)$ rpm		
Viscosity Meters and Cups ^{FO}	1.033 Pa·s	0.005 9 Pa·s	Viscosity Standard Fluids	PR-CAL-025
	43.670 Pa·s	0.011 Pa·s		
	67.810 Pa·s	0.016 Pa·s		
Volumetric Ware/ Equipment ^{FO}	1 mL to 20 mL	$(5 + 0.3 V)$ μ L	Gravimetric method reference to mass balances and OIML E2 weights	PR-CAL-050
	20 mL to 200 mL	$(1.4 + 0.5 V)$ μ L		
	200 mL to 500 mL	$(73 + 0.1 V)$ μ L		
	500 mL to 1 000 mL	$(-36 + 0.3 V)$ μ L		
	1 000 mL to 10 000 mL	$(-708 + 1.0 V)$ μ L		
	10 000 mL to 40 000 mL	$(-35 + 0.9 V)$ μ L		
Hydrometers ^{FO}	0.6 SG to 1.25 SG	0.000 11 SG	Standard Hydrometer	PR-CAL-028
Density meters ^{FO}	0.838 3 g/mL	0.000 17 g/mL	Density Standards	PR-CAL-028
	0.981 3 g/mL			
Absolute, Barometric Pressure Gauges ^{FO}	600 hPa to 1100 hPa	0.69 hPa	Reference Gauge Testo 176 P1	PR-CAL-055



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
IR Spectrometers at the listed wavelengths ^{FO}	539.41 cm x 10 ⁻¹	1.42 cm x 10 ⁻¹	Standard reference filter NIST SRM 1921b	PR-CAL-062
	841.79 cm x 10 ⁻¹	0.72 cm x 10 ⁻¹		
	906.63 cm x 10 ⁻¹	0.22 cm x 10 ⁻¹		
	1 028.27 cm x 10 ⁻¹	0.18 cm x 10 ⁻¹		
	1 069.22 cm x 10 ⁻¹	0.52 cm x 10 ⁻¹		
	1 154.50 cm x 10 ⁻¹	0.12 cm x 10 ⁻¹		
	1 582.98 cm x 10 ⁻¹	0.08 cm x 10 ⁻¹		
	1 601.29 cm x 10 ⁻¹	0.12 cm x 10 ⁻¹		
	1 942.97 cm x 10 ⁻¹	0.66 cm x 10 ⁻¹		
	2 849.48 cm x 10 ⁻¹	0.30 cm x 10 ⁻¹		
	3 001.20 cm x 10 ⁻¹	0.13 cm x 10 ⁻¹		
	3 025.99 cm x 10 ⁻¹	0.32 cm x 10 ⁻¹		
	3 060.16 cm x 10 ⁻¹	0.17 cm x 10 ⁻¹		
3 082.26cm x 10 ⁻¹	0.14 cm x 10 ⁻¹			
Spectrophotometer to measure Absorbance At the wavelengths 250 nm to 635 nm ^{FO}	0.030 Abs	0.002 4 Abs	Neutral Density Filters NIST 2031a	PR-CAL-038
	0.50 Abs	0.004 2 Abs		
	1.0 Abs	0.004 7 Abs		
Spectrophotometer to measure Transmittance At the wavelengths 250 nm to 635 nm ^{FO}	93 T %	0.51 T %		
	31 T %	0.31 T %		
	10 T %	0.11 T %		
Spectrophotometer to output light at fixed point wavelengths ^{FO}	240 nm to 640 nm	0.17 nm	Holmium Oxide Glass	PR-CAL-038



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Measurement Devices ^{FO}	-200 °C to -21 °C	0.031 °C	PRT Thermometer Liquid Bath Dry Block	PR-CAL-029, PR-CAL-030, PR-CAL-031
	-20 °C to 200 °C	0.028 °C		
	200 °C to 420 °C	0.11 °C		
Temperature Measurement "System Accuracy", Oven, Heaters, Incubators, Furnaces, Chambers, Moisture Analyzers ^{FO}	-200 °C to 1 000 °C	1.2 °C	Fluke 724 with Thermocouple	PR-CAL-042, PR-CAL-047
Equipment to Measure and Output Relative Humidity ^{FO}	5 % RH to 95 % RH	1.0 % RH	Vaisala HMP75 Humidity Chamber	PR-CAL-023
IR Thermometers, Pyrometers ^{FO}	30 °C to 60 °C	1.1 % of reading	Blackbody Calibrator with PRT Thermometer	PR-CAL-039
	60 °C to 100 °C	1.2 % of reading		
	100 °C to 500 °C	1.3 % of reading		

Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output Frequency ^{FO}	5 Hz to 99.99 kHz	0.11 % of reading + 0.02 Hz	Fluke 117	PR-CAL-026
	100 Hz to 999 Hz	0.11 % of reading + 0.2 Hz		
	1 kHz to 9.999 kHz	0.11 % of reading + 2 Hz		
	10 kHz to 99.99 kHz	0.11 % of reading + 20 Hz		
Stopwatch ^{FO}	1 hr to 3 hr	0.12 s	Master Stopwatch	PR-CAL-027
	3 hr to 10 hr	0.23 s		
Process timers ^{FO}	300 s to 9 000 s	0.13s		

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.



Certificate of Accreditation: Supplement

BDC Calibration

Av. Gregorio Luperón #51, Los Restauradores,
Santo Domingo 10137, República Dominicana
Contact Name: Mr. Franco Gigliore Phone: 809-338-8888

Accreditation is granted to the facility to perform the following testing:

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term R represents rotational velocity in rpm as appropriate to the uncertainty statement.
7. The term V represents volume in units appropriate to the uncertainty statement.
8. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
9. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.